

# Comparison of Membrane Inlet Mass Spectrometry (MIMS) to Isotope Ratio Mass Spectrometer (IRMS)

**Website:** <https://www.bco-dmo.org/dataset/778065>

**Data Type:** experimental

**Version:** 1

**Version Date:** 2019-10-02

## Project

» [EAGER: Collaborative Research: Detection limit in marine nitrogen fixation measurements - Constraints of rates from the mesopelagic ocean](#) (EAGER NitFix)

Contributors	Affiliation	Role
<a href="#">Granger, Julie</a>	University of Connecticut (UConn)	Principal Investigator
<a href="#">Bourbonnais, Annie</a>	University of Massachusetts Dartmouth (UMass Dartmouth)	Co-Principal Investigator
<a href="#">Wilson, Samuel T.</a>	University of Hawai'i (UH)	Co-Principal Investigator
<a href="#">Biddle, Mathew</a>	Woods Hole Oceanographic Institution (WHOI BCO-DMO)	BCO-DMO Data Manager

## Abstract

In order to ensure that  $^{15}\text{N}_2$  atom% effectuated by Membrane Inlet Mass Spectrometry, we tested whether the mass-specific signal ratios returned by MIMS are equivalent with parallel measurements on a Isotope Ratio Mass Spectrometer, on the basis that the latter returns accurate ratios (see 778000 cal\_iso). MIMS and IRMS returned identical values for parallel samples.

## Table of Contents

- [Dataset Description](#)
  - [Methods & Sampling](#)
  - [Data Processing Description](#)
- [Data Files](#)
- [Related Publications](#)
- [Parameters](#)
- [Instruments](#)
- [Project Information](#)
- [Funding](#)

## Dataset Description

Comparison of MIMS to IRMS

## Methods & Sampling

We tested whether the mass-specific signal ratios returned by MIMS accurately quantitate the relative isotopic abundance of  $^{15}\text{N}_2$  in the range typically used for  $\text{N}_2$  fixation incubations. Incremental  $^{15}\text{N}_2$  gas aliquots were equilibrated in seawater, and respective solutions were then dispensed into replicate gas-tight vials.  $^{15}\text{N}$  atom % of dissolved  $\text{N}_2$  values were measured in parallel by continuous flow IRMS on a GV Isoprime IRMS (Charoenpong et al., 2014) and by MIMS (Bay Instruments), on the assumption that values returned by IRMS are accurate.

## Data Processing Description

BCO-DMO Processing Notes:

- table was extracted from original spreadsheet.

- added conventional header with dataset name, PI name, version date
- modified parameter names to conform with BCO-DMO naming conventions

[ [table of contents](#) | [back to top](#) ]

---

## Data Files

File
<b>mims_2_irms.csv</b> (Comma Separated Values (.csv), 465 bytes) MD5:7b8dbeb0ef6ce3be753fead18af7c9e9 Primary data file for dataset ID 778065

[ [table of contents](#) | [back to top](#) ]

---

## Related Publications

White, A. E., Granger, J., Selden, C., Gradoville, M. R., Potts, L., Bourbonnais, A., Fulweiler, R. W., Knapp, A. N., Mohr, W., Moisaner, P. H., Tobias, C. R., Caffin, M., Wilson, S. T., Benavides, M., Bonnet, S., Mulholland, M. R., & Chang, B. X. (2020). A critical review of the 15N2 tracer method to measure diazotrophic production in pelagic ecosystems. *Limnology and Oceanography: Methods*, 18(4), 129-147. Wiley.  
<https://doi.org/10.1002/lom3.10353>

*Results*

[ [table of contents](#) | [back to top](#) ]

---

## Parameters

Parameter	Description	Units
Subsample	subsample identification	unitless
MIMS_at_pcnt	MIMS at%	unitless
stdev_MIMS_at_pcnt	stdev MIMS at%	unitless
IRMS_15N_at_pcnt	IRMS 15N at%	unitless
stdev_IRMS_15N_at_pcnt	stdev IRMS 15N at%	unitless
IRMS_at_pcnt_no_furnace	IRMS at% no furnace	unitless
stdev_IRMS_at_pcnt_no_furnace	stdev IRMS at% no furnace	unitless

[ [table of contents](#) | [back to top](#) ]

---

## Instruments

<b>Dataset-specific Instrument Name</b>	Isotope Ratio Mass Spectrometer
<b>Generic Instrument Name</b>	Isotope-ratio Mass Spectrometer
<b>Dataset-specific Description</b>	continuous flow Delta V Isotope Ratio Mass Spectrometer (Smith et al. 2015), and continuous flow-GV Isoprime IRMS (Charoenpong et al., 2014)
<b>Generic Instrument Description</b>	The Isotope-ratio Mass Spectrometer is a particular type of mass spectrometer used to measure the relative abundance of isotopes in a given sample (e.g. VG Prism II Isotope Ratio Mass-Spectrometer).

<b>Dataset-specific Instrument Name</b>	Membrane Inlet Mass Spectrometer
<b>Generic Instrument Name</b>	Membrane Inlet Mass Spectrometer
<b>Dataset-specific Description</b>	Membrane Inlet Mass Spectrometer (Bay Instruments)
<b>Generic Instrument Description</b>	Membrane-introduction mass spectrometry (MIMS) is a method of introducing analytes into the mass spectrometer's vacuum chamber via a semipermeable membrane.

[ [table of contents](#) | [back to top](#) ]

## Project Information

### **EAGER: Collaborative Research: Detection limit in marine nitrogen fixation measurements - Constraints of rates from the mesopelagic ocean (EAGER NitFix)**

**Coverage:** North Atlantic Ocean, Pacific Ocean

#### NSF Award Abstract:

The availability of nitrogen is required to support the growth and production of organisms living in the surface of our global ocean. This element can be scarce. To alleviate this scarcity, a special class of bacteria and archaea, called nitrogen fixers, can derive the nitrogen needed for growth from nitrogen gas. This project would carefully examine one specific method for measuring nitrogen fixation that has been used recently to suggest the occurrence of small amounts of nitrogen fixation in subsurface ocean waters. If these reports are verified, then a revision of our understanding of the marine nitrogen cycle may be needed. The Ocean Carbon and Biogeochemistry program will be used as a platform to develop community consensus for best practices in nitrogen fixation measurements and detection of diversity, activity, and abundances of the organisms responsible. In addition, a session will be organized in a future national/international conference to communicate with the broader scientific community while developing these best practices.

The goal of this study is to conduct a thorough examination of potential experimental and analytical errors inherent to the  $^{15}\text{N}_2$ -tracer nitrogen fixation method, in tandem with comprehensive molecular measurements, in mesopelagic ocean waters. Samples will be collected and experimental work conducted on a cruise transect in the North Atlantic Ocean, followed by analytical work in the laboratory. The specific aims of this study are to (1) determine the minimum quantifiable rates of  $^{15}\text{N}_2$  fixation based on incubations of mesopelagic waters via characterization of sources of experimental and analytical error, and (2) seek evidence of presence and expression of nitrogen fixation genes via comprehensive molecular approaches on corresponding samples. The range of detectable rates and diazotroph activity from the measurements made in this study will be informative for the understanding of the importance of nitrogen fixation in the oceanic

nitrogen budget.

[ [table of contents](#) | [back to top](#) ]

---

## Funding

Funding Source	Award
<a href="#">NSF Division of Ocean Sciences (NSF OCE)</a>	<a href="#">OCE-1732246</a>

[ [table of contents](#) | [back to top](#) ]